

Report on MS2 bacteriophage inactivation by medifiber Inc.


Submitted to Medifiber Inc.

By Min Cho, Ph.D.

Safety, Environmental, Life Science (SELS) institute, Chonbuk National University

April 2020


1. Test Period: 2020. 2. 17 ~ 2020. 3. 16
2. Tested membrane modules: fiber #A and fiber #B (CAZ) provided by Medifiber Inc.
3. Test conditions: Triplicates, MS-2 bacteriophage.
4. Results



Fiber #A

MS2 bacteriophage inactivation

Fiber #A : 99.6% (10min)
: 99.998% (30min)



Fiber #B (CAZ)

MS2 bacteriophage inactivation

Fiber #B (CAZ) : 99.8% (1min)
: 99.95% (5min)
: >99.999% (10min)

I certify that the above information is true and correct to the best of my knowledge.

Signature

2020/ April

Date